

LifeScience PLUS, Inc. P.O. BOX 60783 Palo Alto, CA 94306

K071578

510(k) Summary

1. Submitter/Owner: LifeScience PLUS, Inc.

473 Sapena Court, Suite #7

Santa Clara, CA 95054 (Tel) 1-650-565-8172 (Toll free) 1-877-587-5433 (fax) 1-440-445-2752 P.O. BOX 60783 Palo Alto, CA 94306

SEP 2 7 2007

2. Contact Person:

Vicky Feng

3. Date: June 4, 2007

4. Device Name

Trade Name: BloodSTOP and BloodSTOP iX Hemostatic Gauze

Common Name: Hemostatic wound dressing

Product Code: FRO, unclassified

5. Predicate Device: Seal-On Hemostatic Powder Spray, K010933

6. Device Description and Intended Use:

BloodSTOP and BloodSTOP iX Hemostatic Gauze are made from regenerated cotton cellulose, chemically treated to become water-soluble. When contacting blood and exudates, they expand into clear gel, thereby adhering and creating pressure to seal the wound.

Intended Use: OTC topical control of bleeding from minor cuts and abrasions of the skin surface.

7. Substantial Equivalence:

BloodSTOP and BloodSTOP iX Hemostatic Gauze have the same intended use as the predicate device. They are of similar composition as the predicate device, being made from chemically treated plant-source cellulose. They are similar in principle of operation in that both transform into a gel, covering and protecting the wound while hemostasis is achieved. They differ in presentation and method of application, as the predicate is a microdispersed cellulose in aerosol form. Animal studies included in this submission show that BloodSTOP and BloodSTOP iX are equivalent or faster than the predicate Seal-On Powder Spray in time to hemostasis. BloodSTOP and BloodSTOP iX are biocompatible, as demonstrated by test results included in this premarket submission. BloodSTOP and BloodSTOP iX Hemostatic Gauze are substantially equivalent to the cited predicate device in intended use and technological characteristics. Differences in technological characteristics do not raise new issues of safety.



Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

LifeScience PLUS, Inc. % Audrey Vitale 22 Stony Hill Drive Mystic, Connecticut 06355

SEP 2 7 2007

Re: K071578

Trade/Device Name: BloodSTOP and BloodSTOP iX Hemostatic Gauze

Regulatory Class: Unclassified

Product Code: FRO

Dated: September 13, 2007 Received: September 20, 2007

Dear Ms. Audrey Vitale:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally

Page 2 – Ms. Audrey Vitale

marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address http://www.fda.gov/cdrh/industry/support/index.html.

Sincerely yours

Mark N. Melkerson

Director

Division of General, Restorative and Neurological Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

3. Indications for Use

510(k) Number (if known): K071578
Device Name: BloodSTOP and BloodSTOP iX Hemostatic Gauze
Indications For Use:
Topical control of bleeding from minor cuts and abrasions of the skin surface
(Division Sign-Off) Division of General, Restorative, and Neurological Devices 510(k) Number
Prescription Use AND/OR Over-The-Counter Use (Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)
(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)
Concurrence of CDRH, Office of Device Evaluation (ODE)